



MENU

Results for EFFECTS OF AM... >

Effects of Ambient Acoustic Noise on Auditory Brainstem Response to Level-...



Effects of Ambient Acoustic Noise on Auditory Brainstem Response to Level-Specific Chirp and Click Stimuli in Normal-Hearing Adults

By

Dzulkarnain, AAA (Dzulkarnain, Ahmad Aidil Arafat) ; Rahed, BAM (Rahed, Balqis Aqilah Mat) ; Shahrudin, FA (Shahrudin, Fatin Amira) ; Jamal, FN (Jamal, Fatin Nabilah) ; Zakaria, MN (Zakaria, Mohd Normani)

[View Web of Science ResearcherID and ORCID](#) (provided by Clarivate)

Source

JOURNAL OF AUDIOLOGY AND OTOTOLOGY
DOI: 10.7874/jao.2021.00689

Early Access

MAY 2022

Indexed

2022-06-11

Document Type

Article; Early Access

Abstract

Background and Objectives: Despite few reports on the influence of ambient acoustic noise on auditory brainstem response (ABR) to click stimuli, its effects on ABR to level-specific (LS) stimuli have not been systematically investigated. This study aimed to investigate the influence of ambient acoustic noise on ABR findings using both LS chirp and click stimuli. **Subjects and Methods:** Twelve normal-hearing adults participated in this repeated measure design study. The ABRs were acquired at 80,

50, and 30 dBnHL using two stimuli (LS chirp and click) under two conditions (quiet and noise). The ABRs under noise conditions were acquired using babble noise and white noise. The noise level was set at 55 dBA. Two-way repeated measure analysis was used to identify the main effects of the test conditions, stimulus types, and their interactions at a 95% confidence level. Results: No significant influence of ambient acoustic noise on ABR findings was identified at all intensity levels. No significant difference was found in the number of signal averages to reach the 0.04 μ V residual noise as stopping criteria among the ABRs recorded with different types of stimuli and test conditions. The ABR waves I and V amplitudes were larger with LS chirp than with click stimulus. Conclusions: Ambient acoustic noise has no significant influence on ABR findings and the ABR test time based on the 55 dBA noise level used in this study.

Keywords

Author Keywords: Acoustics; Noise; Evoked potentials; Auditory; Brain stem

Keywords Plus: ABR

Addresses

¹ Int Islamic Univ Malaysia, Dept Audiol & Speech Language Pathol, Kulliyah Allied Hlth Sci, Jalan Sultan Ahmad Shah, Kuantan 25200, Pahang, Malaysia

² Univ Sains Malaysia, Sch Hlth Sci, Audiol Programme, Kubang Kerian, Kelantan, Malaysia

**Categories/
Classification**

Research Areas: Otorhinolaryngology

Citation [1 Clinical & Life](#) , [1.150](#) , [1.150.421 Auditory](#)
Topics: [Sciences](#) [Hearing Loss](#) [Disorders](#)

Sustainable Development Goals: [03 Good Health and Well-being](#)

**Web of Science
Categories**

Otorhinolaryngology

[+ See more data fields](#)

Citation Network

In Web of Science Core Collection

Use in Web of Science

1

4