

Documents

Aziz, A.A.^a, Azmi, M.A.M.^a, Zulkifli, M.N.^a, Nordin, A.N.^b, Bahrain, A.K.^a, Makmon, F.Z.^a, Badruzzaman, N.A.^a, Sabdin, S.^a

Rapid fabrication and characterization of PDMS microfluidics device using printed conductive silver ink
(2019) *Materials Today: Proceedings*, 16, pp. 1661-1667.

DOI: 10.1016/j.matpr.2019.06.032

^a Electronic Technology Department, Universiti of Kuala Lumpur British Malaysia Institute (UniKL-BMI), Jalan Sungai Pusu, Gombak, Selangor Darul Ehsan, 53100, Malaysia

^b Electrical and Computer Engineering Department, Engineering Faculty International Islamic University Malaysia, Kuala Lumpur, Malaysia

Abstract

Currently, new fabrication methods for rapid prototyping of microfluidic devices has bring attention among researchers from various fields. The main reason is the use of lithographic equipment in standard fabrication method for master fabrication require trained personnel and high cost disposable materials. In this paper, a new fabrication method via printed conductive silver ink is introduced in order to minimize the current fabrication process through elimination of the photolithography process. As a result, a usable, faster and cost effective prototyping process of a microfluidic based biosensor devices able to be produced. In this method, the conductive silver ink was used to create a master template contains microchannel onto glass-reinforced epoxy laminate sheets (FR-4). Then, Polydimethylsiloxane (PDMS) were used to replicate the structure. Next, the digital microscope was used to obtain images to demonstrate geometry fluidics structure with channel dimension up to 200 μm width. Moreover, a 3D profilometer was used to evaluate their thickness and roughness of the microfluidics structure. The stability test demonstrates the consistency of flow sequence inside the serpentine and separation channel. © 2019 Elsevier Ltd. All rights reserved.

Author Keywords

Microfluidic; PDMS; Rapid prototyping; Silver ink; Softlithography

Funding details

FRGS2016-1

British Malaysian Institute, Universiti Kuala LumpurBMI, UniKL

Correspondence Address

Aziz A.A.; Electronic Technology Department, Universiti of Kuala Lumpur British Malaysia Institute (UniKL-BMI), Jalan Sungai Pusu, Malaysia; email: azmanaziz@unikl.edu.my

Editors: Mohammad N.M., Muhamad N., Ismail H.

Publisher: Elsevier Ltd

Conference name: 2017 Conference on Biomedical and Advanced Materials, Bio-CAM 2017

Conference date: 28 November 2017 through 29 November 2017

Conference code: 151551

ISSN: 22147853

Language of Original Document: English

Abbreviated Source Title: Mater. Today Proc.

2-s2.0-85072393528

Document Type: Conference Paper

Publication Stage: Final

Source: Scopus