



Document details

< Back to results | 1 of 1

↗ Export ↴ Download 🖨 Print ✉ E-mail 💾 Save to PDF ☆ Add to List More... >

View at Publisher

Journal of Electronic Imaging
Volume 28, Issue 2, 1 March 2019, Article number 021002

In-socket sensory system with an adaptive neuro-based fuzzy inference system for active transfemoral prosthetic legs (Article)

Mohd Yusof, N.H.^a, Hamzaid, N.A.^a, Jasni, F.^{a,b}, Lai, K.W.^a ✉ 👤

^aUniversity Malaya, Department of Biomedical Engineering, Faculty of Engineering, Kuala Lumpur, Malaysia

^bInternational Islamic University Malaysia, Department of Mechatronics Engineering, Faculty of Engineering, Kuala Lumpur, Malaysia

Abstract

An in-socket sensory system enables the monitoring of transfemoral amputee movement for a microprocessor-controlled prosthetic leg. User movement recognition from an in-socket sensor allows a powered prosthetic leg to actively mimic healthy ambulation, thereby reducing an amputee's metabolic energy consumption. This study established an adaptive neurofuzzy inference system (ANFIS)-based control input framework from an in-socket sensor signal for gait phase classification to derive user intention as read by in-socket sensor arrays. Particular gait phase recognition was mapped with the cadence and torque control output of a knee joint actuator. The control input framework was validated with 30 experimental gait samples of the in-socket sensory signal of a transfemoral amputee walking at fluctuating speeds of 0 to 2 km · h⁻¹. The physical simulation of the controller presented a realistic simulation of the actuated knee joint in terms of a knee mechanism with 95% to 99% accuracy of knee cadence and 80% to 90% accuracy of torque compared with those of normal gait. The ANFIS system successfully detected the seven gait phases based on the amputee's in-socket sensor signals and assigned accurate knee joint torque and cadence values as output. © 2018 SPIE and IS&T.

SciVal Topic Prominence ⓘ

Topic: Prosthetics | Artificial limbs | Prosthetic leg

Prominence percentile: 97.245 ⓘ

Author keywords

adaptive neurofuzzy inference system controller in-socket sensory system prosthetic knee joint transfemoral leg

Indexed keywords

Engineering controlled terms:

Adaptive control systems Artificial limbs Controllers Energy utilization Joint prostheses Joints (anatomy) Motion estimation Pattern recognition

Engineering uncontrolled terms

Adaptive neuro-fuzzy inference system Fuzzy inference systems Gait phase classifications Microprocessor controlled Prosthetic knee joint Realistic simulation Sensory system Transfemoral

Engineering main heading:

Fuzzy inference

Metrics ⓘ View all metrics >

Field-Weighted
Citation Impact



PlumX Metrics



Usage, Captures, Mentions,
Social Media and Citations
beyond Scopus.

Cited by 0 documents

Inform me when this document
is cited in Scopus:

Set citation alert >

Set citation feed >

Related documents

Find more related documents in
Scopus based on:

Authors > Keywords >

Funding details


Funding sponsor	Funding number	Acronym
	PG125-2015B	
Ministry of Higher Education	FP047-2014B	MOHE

Funding text

The authors thank Ms Nur Nadia Asyikin Bt Zakaria for her assistance in the gait illustration. This research was fully supported by University Malaya through Postgraduate Research Fund (PPP) Project No: PG125-2015B and Ministry of Higher Education through Fundamental Research Grant Scheme Grant Project No: FP047-2014B.

ISSN: 10179909
CODEN: JEIME
Source Type: Journal
Original language: English

DOI: 10.1117/1.JEI.28.2.021002
Document Type: Article
Publisher: SPIE

 Lai, K.W.; University Malaya, Department of Biomedical Engineering, Faculty of Engineering, Kuala Lumpur, Malaysia; email:lai.khinwee@um.edu.my

© Copyright 2019 Elsevier B.V., All rights reserved.

About Scopus

What is Scopus

Content coverage

Scopus blog

Scopus API

Privacy matters

Language

日本語に切り替える

切换到简体中文

切换到繁體中文

Русский язык

Customer Service

Help

Contact us